

Chlorine**Code : XL-506****Range : 0.1, 0.2, 0.4, 0.7, 1 & 2 ppm as Chlorine.****AQUA-XL****Water Analysing
Kits****Directions for use :**

1. Add 2 drops of **Chlorine Reagent** in the Test jar.
2. Pour Water Sample upto 10 ml mark and mix well. Wait for 2 minutes.
3. Now rinse the empty comparator tube with above solution and fill upto the half level of the comparator tube.
4. Insert this tube in one of the sample slot (S) of the Comparator in such a way that the sample tube colour is close to the matching colour of the comparator.
5. Note down the **ppm level of Chlorine as CL₂ after colour comparison.**

Chlorine**Code****: XL-516****Range : 0.5, 1, 2, 3, 4 & 5 ppm as Chlorine..****AQUA-XL****Water Analysing Kits****Directions for use :**

1. Add 2 drops of **Chlorine Reagent** in the Test jar.
2. Pour Water Sample upto 10 ml mark and mix well. Wait for 2 minutes.
3. Now rinse the empty comparator tube with above solution and fill upto the half level of the comparator tube.
4. Insert this tube in one of the sample slot (S) of the Comparator in such a way that the sample tube colour is close to the matching colour of the comparator.
5. Note down the **ppm level of Chlorine as CL₂ after colour comparison.**